





	装 置 名	重量 (kg)	発熱量 (w)	装置寸法(W×D×H)mm
21	SOMATOM Sensation 64: Gantry	2000	1380	2271× 940×1983
22	天井ケーブルレール	200		4715× 893× 288
23	パワーディストリビューター	550	1200	900× 750×1800
24	1RS3キャビネット	120	1600	600× 745×1800
25	熱交換機 Heat exchanger (水冷)	200	500	900× 750×1800
26	操作コンソール: DMC (181TF)	60	70	1400× 800× 700
27	IES+ICS+コンテナー	58	700	660× 858× 700
28	移動制御キャビネット	58	700	660× 300×1000
29	床ガントリースライド用モーターユニット	120		300× 3100× 45
30	ケアビジョン	58		

冷水設備 屋外に設置 (参考)



- 免責事項**

  - ・本図面に記載されたデータや装置の仕様は予告無く変更場合があります。
  - ・本図面はツメンス社製装置の正常な稼働を目的としたものです。

既設装置番号：WBS-0004775.01~02

変更内容				日程	担当	実施	変更内容				日程	担当	実施	お客様ご来館印	期 日	2014.08.13 Watanabe	静岡県立こども病院 殿	4F 撮影室	
01. 新規 (OR-TAへ変更)	2012.07.17	渡辺	津山	06. 前室側扉変更、空調・37度追加など	2014.05.26	渡辺	津山	02. 可動範囲壁追加	2013.06.27	渡辺	津山	07. 周辺設備追加	2014.08.13						渡辺
03. 可動範囲壁面へ前室追加、天井長尺追加、HEPA追加、一部壁変更	2013.07.11	渡辺	津山	08.															
04. 一部壁変更、補修追加、操作室側扉変更など	2013.09.30	渡辺	津山	09.															
05. 資料室一部壁変更	2014.02.20	渡辺	津山	10.															


